

Message

From: Hanchett, James (DPH) [/O=COMMONWEALTH OF MASSACHUSETTS/OU=MASSMAIL-01/CN=RECIPIENTS/CN=JAMES.HANCHETT]
Sent: 1/9/2013 2:14:32 PM
To: Brooks, Nancy (POL) [/O=COMMONWEALTH OF MASSACHUSETTS/OU=EXCHANGE ADMINISTRATIVE GROUP (FYDIBOHF23SPDLT)/CN=RECIPIENTS/CN=Nancy.wong047]
Subject: Amherst FTIR Update

Hi Nancy,

I replaced the FTIR power supply you sent us from the Jamaica Plain Lab. Unfortunately the same error message continues to appear "Replace Laser". Something else must be wrong, so a service call will be necessary when funding is available. We can get buy now with our mass specs. The only time IR was necessary was cocaine free base confirmation for Federal Courts and a GC or MS of the sample in Petroleum Ether will work.

Thanks,

Jim Hanchett
Department of State Police
Amherst Drug Lab
Room N251 Morrill I
637 North Pleasant Street
Amherst, MA 01003
Phone 413-545-2607
Fax 413-545-2608
Cell [REDACTED]

The preceding email message (including any attachments) contains information that may be confidential, may be protected by the attorney-client or other applicable privileges, or may constitute non-public information. It is intended to be conveyed only to the designated recipient(s) named above. If you are not an intended recipient of this message, please notify the sender by replying to this message and then delete all copies of it from your computer system. Any use, dissemination, distribution, or reproduction of this message by unintended recipients is not authorized and may be unlawful.